

Declaration of Conformity to EU Medical Device Regulation 2017/745

| | |
|--|--|
| Legal Manufacturer | Coloplast A/S Holtedam 1, 3050 Humlebaek, DK SRN: DK-MF-000025526 |
| EU Product Classification according to Annex VIII | I Rule Number: 1 |
| Intended Purpose | Uro Connector is intended to ensure connection between an urostomy bag and a drainage night bag. |
| Basic UDI-DI | 57089329102477A |
| Conformity to Common Specification(s) | No relevant Common Specification to list |
| Conformity to other Union Legislation(s) | No relevant Union Legislation to list |

This EU Declaration of Conformity is applicable for following catalogue numbers:

| Catalogue Number | Product Name | Original CE Marking Date yyyy-mm-dd |
|-------------------------|--------------------------------|--|
| 09800 / 098000 / 9800 | Assura / Alterna Uro connector | 2010-10-13 |

This EU Declaration of Conformity is issued under the sole responsibility of Coloplast A/S. Coloplast A/S declares that the devices covered by the present declaration are in conformity with the Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices and the requirements specified herein are fulfilled.

Date of signature: 2023-08-25
yyyy-mm-dd

Place of signature: Humlebaek, Denmark
Place, Country

DKBENB, Benedikte Blom, Head of Regulatory Affairs

Signed on behalf of Coloplast A/S:



Name, Title